

MEDIVATORS SSD102

DATA LOG Machine ID: 6012
19 Oct 2016 15:42 Page 01
DATE Time SP EVENT
001 19 Oct 15:00 41 Start 000000235
002 Score 000000032
003 Operator 000000527
004 Patient 000000005
005 Physician
006 19 Oct 15:00 41 Log Pass 00000
007 19 Oct 15:00 41 Flush 00000
008 Cycle Count 206
009 19 Oct 15:00 41 Disinfect 00000
010 19 Oct 15:00 41 Temp 00000

SSD Active

Flush Disinfect

Rinse 1 Rinse 2

Air Alcohol

STATUS

WARNING

1 2 3

4 5 6

7 8 9

* 0 #

ID Data

Program

Add Air

Heater On

Disinfectant Dump

Set Up

ENTER

CANCEL

Start

Stop



Endoscope reprocessing the way it should be™

MEDIVATORS®



Dimensions

46 H x 22 W x 21 D (inches)
 117 H x 56 W x 53 D (cm)
 Height with lid open: 65 inches (165 cm)

Weight

220 lbs (100 kg)

Electrical Requirements

110VAC, 60Hz, 8 amp (North America)
 230VAC, 50/60Hz, 4 amp (International)

Electrical Safety Certifications

ETL Intertek Listed #75811
 UL Standard 61010
 CSA/CSA Standard C22.2, No. 601.1

Reservoir Capacities

Disinfectant: 4 gallons (15.1 liters)
 Alcohol and detergent: 27oz each (800 ml)

Disinfectant Temperature Range

Room temp to 125°F (52°C)

SSD Ordering Information

Model number	Description
SSD-102LT	SSD Endoscope Reprocessor with Leak Tester

SSD Accessory Ordering Information


Part number	Description
ML02-0059	Rapicide High Level Disinfectant and Sterilant
ML02-0120	Rapicide Potency Indicator Test Strips
MD01-1002	SSD Mobile Dolly

Data Management Packages

DSD-2021	Designed for SSD Software Version 5.07
DSD-2022	Designed for SSD Software Version 4.11 or Earlier

Vapor Management Options

MK01-0055	(110V) Active Vapor Management System
MK01-0056	(230V) Active Vapor Management System
DSD-1070R	Passive Vapor Management System



Minntech Corporation
 14605 28th Avenue North
 Minneapolis MN 55447
 Tel: +1-763-553-3300
 Fax: +1-763-553-3387

Minntech BV
 Sourethweg 11
 6422 PC Heerlen
 The Netherlands
 Tel: +31-45-5-471-471
 Fax: +31-45-5-429-695

Minntech Asia/Pacific Pte Ltd
 1 A International Business Park
 #05-01 Singapore 609933
 Tel: +65-6227-9698
 Fax: +65-6225-6848

Minntech Beijing Representative Office
 Room 708, 7th Floor Kaiheng Center, Block B
 No. 2 Chaoyangmennei Street,
 Dongcheng District, Beijing China, 100010
 Tel: +8610-6567-8446
 Fax: +8610-6567-8445

For additional information please contact your Medivators representative
 or visit our website at www.minntech.com/medivators